

NOTE

## EFFECT OF OXYGEN ELECTRODE RESPONSE ON $K_L a$ MEASUREMENT

Ho Nam CHANG\*, Dong Jin KIM, and Murray MOO-YOUNG\*\*

Department of Chemical Engineering, Korea Advanced Institute of Science and Technology,  
P.O. Box 131, Daejeon, Seoul, Korea

\*\*Department of Chemical Engineering, University of Waterloo, Waterloo, Ontario N2L 3G1, Canada

(Received 22 February 1989 • accepted 19 June 1989)

**Abstract**—A study on the effect of oxygen electrode response on  $K_L a$  measurement was carried out by simulation. As oxygen electrode response time increases, the estimated  $K_L a$  decreases when we did not consider the electrode response time. And  $K_L a$  estimation by slope method is closer to real one than curve fitting method.

### INTRODUCTION

The volumetric mass transfer coefficient ( $K_L a$ ) in gas-liquid dispersions has always played a central role in the design and characterization of bioreactor for aerobic processes. It determines the amount of oxygen that can be supplied to the liquid phase where chemical and biochemical reactions requiring oxygen are going on. The coefficient is a function of the agitation speed, gas flow rate, geometrical properties of the vessel and the agitator and, finally, of physicochemical properties of the liquid. Although some relationships between the important variables are already known, the overall process is very complex and it is not possible to obtain precise quantitative values for the coefficient of the given reactor. By far the most widely used and accepted method is the dynamic method[1], which requires the use of a dissolved oxygen electrode. The experimental procedure is reasonably straight forward. A step change is established in the oxygen concentration of the inlet gas phase and the response of the dissolved oxygen electrode is recorded. With an appropriate mathematical description of the physicochemical phenomena of the experiment, we can estimate  $K_L a$  from the experimental data.

### THEORY

When we consider both gas phase and liquid phase dynamics for oxygen transfer in a continuous completely mixed tank fermentor where microbial bio-

mass consumes a constant amount of oxygen per unit time, the governing mass balance equations based on the liquid volume are [2]

$$\frac{dC_g}{dt} = \frac{1}{t_g} (C_{g1} - C_g) - K_L a \frac{1-\epsilon}{\epsilon} (C_L^* - C_L) \quad (1)$$

$$\frac{dC_L}{dt} = K_L a (C_L^* - C_L) - R + D (C_{L1} - C_L) \quad (2)$$

where  $t_g$  is the average gas residence time which is given as  $t_g = (V_L/Q_g) \cdot [\epsilon/(1-\epsilon)]$ , and  $R$  represents the microbial oxygen uptake rate.

If the dimensionless variables  $X$ ,  $Y$ ,  $\lambda$  and  $t^*$  are defined as [3]

$$X = \frac{C_g - C_g^1}{C_{g1}^1 - C_{g1}^1}, \quad Y = \frac{C_L - C_L^1}{C_L^{1*} - C_L^1} \quad (3)$$

$$\lambda (C_g^1 - C_g^1) = (C_L^{1*} - C_L^1) \quad (4)$$

$$t^* = t/t_g \quad (5)$$

where  $\lambda$  is dimensionless solubility of oxygen, equations (1) and (2) can be rewritten in dimensionless forms:

$$\frac{dX}{dt^*} = 1 - X - a_1 (X - Y) \quad (6)$$

$$\frac{dY}{dt^*} = a_2 (X - Y) - a_3 Y \quad (7)$$

$$X=0, Y=0 \quad \text{at } t^*=0, \quad (8)$$

$$a_1 = K_L a t_g \lambda \frac{1-\epsilon}{\epsilon}, \quad a_2 = K_L a t_g, \quad a_3 = D t_g \quad (9)$$

Then the liquid phase oxygen concentration is given by[4]

\*To whom all correspondence should be addressed.

$$Y(t) = Y(\infty) [1 - \alpha \exp(m_1 t^*) - (1 - \alpha) \exp(m_2 t^*)] \quad (10)$$

$$m_1 = \frac{-q_1 + \sqrt{q_1^2 + 4q_2}}{2}, \quad m_2 = \frac{-q_1 - \sqrt{q_1^2 + 4q_2}}{2} \quad (11)$$

$$q_1 = 1 + a_1 + a_2 + a_3, \quad q_2 = a_2 + a_1 a_3 + a_3 \quad (12)$$

$$\alpha = -\frac{m_2}{m_1 - m_2} \quad (13)$$

Usually  $-m_2$  is at least one order of magnitude larger than  $-m_1$ . For  $t^*$  values greater than 5, only the first exponential term is significant so that

$$Y(t^*) = Y(\infty) [1 - \alpha \exp(m_1 t^*)] \quad (14)$$

Therefore, a semilogarithmic plot of  $1 - Y(t^*)/Y(\infty)$  yields a straight line with a slope of  $m_1$  when  $t^*$  becomes large. From  $m_1$  we can deduce  $K_L a$

$$K_L a t_e = \frac{-m_1(a_3 + 1) - (m_1^2 + a_3)}{(1 + \frac{1-\epsilon}{\epsilon} \lambda a_3) + m_1(1 + \frac{1-\epsilon}{\epsilon} \lambda)} \quad (15)$$

when  $a_3 = 0$ , equation (15) reduces to

$$K_L a t_e = \frac{m_1 - m_1^2}{1 + m_1(1 + \frac{1-\epsilon}{\epsilon} \lambda)} \quad (16)$$

If we employ an oxygen electrode to measure dissolved oxygen concentration, the oxygen electrode response can be obtained by solving a related partial differential equations[5-8]. But in practice a simple first order response with some time lag has been known to adequately describe the oxygen electrode[9,10]. Then the slope obtained by equation (15) will be modified by[4,11]

$$t_e \frac{d\bar{Y}}{dt^*} + \bar{Y} = Y(t^*), \quad \text{at } t^* = 0, \quad \bar{Y}(t^*) = 0 \quad (17)$$

where  $t_e$  is the first order time constant of the electrode. Then from equation (6), (7) and (17), the solution is

$$\begin{aligned} \bar{Y}(t^*) = Y(\infty) \left\{ 1 - \exp\left(-\frac{t^*}{t_e}\right) + \frac{\alpha}{1 + m_1 t_e} \right. \\ \left. \left[ \exp\left(-\frac{t^*}{t_e}\right) - \exp(m_1 t^*) \right] + \frac{1 - \alpha}{1 + m_1 t_e} \right. \\ \left. \left[ \exp\left(-\frac{t^*}{t_e}\right) - \exp(m_2 t^*) \right] \right\} \quad (18) \end{aligned}$$

If we take the slope of the semilogarithmic plot of  $1 - \bar{Y}(t^*)/\bar{Y}(\infty)$  by neglecting the last term, we have

$$\frac{d \ln \left( 1 - \frac{\bar{Y}(t^*)}{\bar{Y}(\infty)} \right)}{dt^*} =$$


$$\begin{aligned} -\frac{1}{t_e} \exp\left(-\frac{t^*}{t_e}\right) \left( 1 - \frac{\alpha}{1 + m_1 t_e} \right) + \frac{\alpha m_1}{1 + m_1 t_e} \exp(m_1 t^*) \\ \exp\left(-\frac{t^*}{t_e}\right) \left( 1 - \frac{\alpha}{1 + m_1 t_e} \right) + \frac{\alpha}{1 + m_1 t_e} \exp(m_1 t^*) \quad (19) \end{aligned}$$

$$\text{slope} = \frac{m_1 - \frac{1}{t_e} \left( \frac{1 + m_1 t_e}{\alpha} - 1 \right) \exp\left[ \left( -\frac{1}{t_e} - m_1 \right) t^* \right]}{1 + \left( \frac{1 + m_1 t_e}{\alpha} - 1 \right) \exp\left[ \left( -m_1 - \frac{1}{t_e} \right) t^* \right]} \quad (20)$$

When  $|m_1|$  is less than  $1/t_e$ , the slope will approach  $m_1$  as  $t^*$  goes to infinite. Thus theoretically  $K_L a$  values from the slope will be less sensitive to the electrode response time than that from the curve fitting of the response curve in the full time domain.

## RESULTS AND DISCUSSION

In Figure 1 the electrode response curves are given with varying characteristic time from equation (18) at the simulation conditions of Table 1. The data in Table 1 are obtained from the results of Chang et al.[12]. The characteristic time is defined as the electrode response time normalized with respect to the inverse of  $K_L a$  to be measured. Thus  $CT=0$  means that the



**Fig. 1. Effect of time constant of dissolved oxygen electrode on the response.**

(A1)  $U_{sg} = 1.30$ ,  $\epsilon = 0.039$ ,  $K_L a = 0.0108$ ,  $t_g = 2.87$   
 (B1)  $U_{sg} = 5.74$ ,  $\epsilon = 0.135$ ,  $K_L a = 0.0388$ ,  $t_g = 2.50$   
 From the top  $CT = 0, 0.2, 0.4, 0.6$  and  $0.8$ , respectively. (A2), (B2): semilogarithmic plot of (A1) and (B1), respectively.

**Table 1. Simulation conditions (D=0)**

$U_{sg}$	$\epsilon$	$K_L a$	$t_g$
1.30	0.039	0.0106	2.87
5.74	0.135	0.0388	2.50

**Table 2. Effect of oxygen electrode response on  $K_L a$**

CT	0	0.2	0.4	0.6	0.8	TI
$K_L a = 0.0106$						
A	1.0	0.615	0.421	0.317	0.254	(0-28)
B	1.0	0.939	0.736	0.578	0.471	(18-28)
C	1.0	0.998	0.937	0.817	0.701	(40-65)
$K_L a = 0.0388$						
A	1.0	0.762	0.618	0.515	0.440	(0-28)
B	1.0	1.0	0.999	0.972	0.884	(18-28)
C	1.0	1.0	1.0	0.984	0.926	(40-65)

All estimated  $K_L a$  values were normalized with respect to those of real  $K_L a$ .

response is free of any distortion due to the electrode. As the characteristic time increases, the saturation of dissolved oxygen electrode response becomes slower and slower (Figure 1a). In Figure 1b the semilogarithmic plot of  $(1-\bar{Y})$  versus time of Figure 1a are given. Because of the presence of the first order lag response, the slope decreases as CT increases. Also the slopes depend on the time interval chosen for the measurement as predicted by equation (20).

In real fermentation systems, dissolved oxygen electrode has a large response time constant and its time constant varies with the number of autoclaving. So we can not know the electrode response time exactly and the estimated  $K_L a$  value using equation (20) is not precise because  $t_e$  in equation (20) is not precise. In this study we assume the time constant as zero and estimate the  $K_L a$  value using equation (16) from the slope of dissolved oxygen electrode response profile at the varying characteristic time and time interval to study the effects of characteristic time and the time interval on  $K_L a$  value.

Table 2 shows that two real  $K_L a$  values of 0.0106 and 0.0388 can be altered depending on the estimation method and the oxygen electrode response time constant and the time interval(TI). The values in Table 2 are the ratios of estimated  $K_L a$  to real  $K_L a$ . The estimated  $K_L a$  values are obtained from the profiles of Figure 1 using equation (14) or (16). Either the increase of  $K_L a$  or  $t_e$  can increase CT which is defined as the multiplication of  $K_L a$  and  $t_e$ , and that decreases the estimated  $K_L a$  value. So the estimated  $K_L a$  will be smaller than the real one when  $CT > 0$ . As a method of

$K_L a$  estimation, the curve fitting method(A) using equation (14) yields  $K_L a$  which are more sensitive to the CT than the slope method. When the slope method (B, C) using equation(16) is employed,  $K_L a$  is less sensitive to the CT than the curve fitting (Powell's) method[13] and it closes to the real one at the later TI. That is shown in a column of Table 2. The values of B and C in a column are greater than that of A when  $CT \neq 0$ . This shows the feasibility of slope method when we estimate  $K_L a$  without knowing  $t_e$ . The value of C is greater than that of B in a column. This means that the later time interval has less electrode response time effect than the earlier time interval. This effect can be seen in equation (20). However, the theoretical prediction applies only to the results for a late time.

In practice, the influence of the electrode is unavoidable, it is severer when  $K_L a$  is larger, and as a result a smaller  $K_L a$  is obtained unless a proper correction is made with respect to the time constant. However, error can be minimized by using an electrode with smaller time constant and also by taking the slope at the time interval with later TI.

Conclusively, the increase of dissolved oxygen electrode response time decreases the estimated  $K_L a$  when we do not consider the response time effect. And  $K_L a$  estimation by slope method at later TI closes to real one than curve fitting method.

**NOMENCLATURE**

- $a_1, a_2, a_3$  : dimensionless coefficient defined in equation (6)
- $C_g$  : gas phase oxygen concentration [mol/L]
- $C_g^1$  : gas phase oxygen concentration of state 1 [mol/L]
- $C_{gi}$  : gas phase oxygen concentration at inlet [mol/L]
- $C_{gi}^1, C_{gi}^2$  : gas phase oxygen concentration at inlet of state 1 and 2 [mol/L]
- $C_L$  : liquid phase oxygen concentration [mol/L]
- $C_L^1$  : liquid phase oxygen concentration of state 1 [mol/L]
- $C_{Li}$  : liquid phase oxygen concentration at inlet [mol/L]
- $C_L^*$  : liquid phase oxygen concentration in equilibrium with gas phase [mol/L]
- $C_L^{*1}, C_L^{*2}$  : liquid phase oxygen concentration in equilibrium with gas phase of state 1 and 2 [mol/L]
- CT : dimensionless time constant of an oxygen electrode [ $t_e K_L a$ ]
- D : dilution rate [1/s]
- $K_L a$  : volumetric mass transfer coefficient [1/s]

$m_1, m_2$	: dimensionless coefficient defined in equation (8)
$Q_g$	: gas flow rate [L/s]
$q_1, q_2$	: dimensionless coefficient defined in equation (9)
$R$	: microbial oxygen uptake rate [mol $O_2$ /L·s]
$t$	: time [s]
$t_e$	: first order time constant of oxygen electrode [s]
$t_g$	: gas residence time [s]
$t^*$	: dimensionless time [ $t/t_g$ ]
$TI$	: dimensionless time interval used to evaluate a slope
$U_{sg}$	: superficial gas velocity [cm/s]
$V_L$	: total fermentor volume [L]
$X$	: dimensionless gas phase oxygen concentration $[(C_g - C_g^1)/(C_{gr}^2 - C_{gd}^1)]$
$Y$	: dimensionless liquid phase oxygen concentration $[(C_L - C_L^1)/(C_L^{2*} - C_L^{1*})]$
$\bar{Y}$	: dimensionless liquid phase oxygen concentration measured by dissolved oxygen electrode

#### Greek Letters

$\alpha$	: dimensionless coefficient defined in equation (10)
$\epsilon$	: gas holdup
$\lambda$	: dimensionless oxygen solubility

#### REFERENCES

1. Bandyopadhyay, B., Humphrey, A.E., and Taguchi, T.: *Biotechnol. Bioeng.*, **9**, 533 (1967).
2. Andre, G., Moo-Young, M., and Robinson, C.W.: *Biotechnol. Bioeng.*, **23**, 1611 (1981).
3. Kim, D.J. and Chang, H.N.: *J. Chem. Tech. Biotechnol.*, **45**, 39 (1989).
4. Kim, D.J.: MS Thesis, KAIST, Seoul, Korea (1987).
5. Heineken, F.G.: *Biotechnol. Bioeng.*, **12**, 145 (1970).
6. Heineken, F.G.: *Biotechnol. Bioeng.*, **13**, 599 (1971).
7. Linek, V.: *Biotechnol. Bioeng.*, **14**, 285 (1972).
8. Wernau, W.C. and Wilke, C.R.: *Biotechnol. Bioeng.*, **15**, 571 (1973).
9. Aiba, S. and Huang, S.Y.: *Chem. Eng. Sci.*, **24**, 1149 (1969).
10. Vadar, F. and Lilly, M.D.: *Biotechnol. Bioeng.*, **24**, 1711 (1982).
11. Chang, H.N.: "Biochemical Engineering", Panghan Press, Seoul (1988).
12. Chang, H.N., Halard, B., and Moo-Young, M.: *Biotechnol. Bioeng.*, in press.
13. Kluester, J.L. and Mize, J.H.: "Optimization Techniques with Fortran", McGraw-Hill, New York, NY (1973).